

CERTIFICATE OF ANALYSIS

Buffer Solution

7.00 pH

Item code:	DC107022 (05/1/107-NR)
Batch code:	12047
Exp. Date:	07/2019
Manufactured:	07/2016
Tolerance:	+/- 0.015 pH
Nominal pH-value:	7.00 pH (at 25°C)
Actual pH-value:	7.00 pH (at 25°C)
Secondary Ref. Material:	Phosphate
Ref. Material LOT-No.:	M6880S01/14

Standardization method:

Actual pH-value as stated above was obtained using a combination glass electrode according to DIN 19268.

Calibration is traceable to national and international standards, PBT and NIST validated by key comparisons.