

CERTIFICATE OF ANALYSIS

Buffer Solution

4.01 pH

Item code:	DC104004 (05/1/107-NR)
Batch code:	11600
Exp. Date:	11/2017
Manufactured:	10/2015
Tolerance:	+/- 0.015 pH
Nominal pH-value:	4.01 pH (at 25°C)
Actual pH-value:	4.01 pH (at 25°C)
Secondary Ref. Material:	Potassiumhydrogenphthalate
Ref. Material LOT-No.:	M4008S01/14

Standardization method:

Actual pH-value as stated above was obtained using a combination glass electrode according to DIN 19268.

Calibration is traceable to national and international standards, PBT and NIST validated by key comparisons.