

Water-i.d. / Pool-i.d. GmbH Daimlerstr. 20 D-76344 Eggenstein Germany www.water-id.com

CERTIFICATE OF ANALYSIS

Buffer Solution 9.18 pH

Item code: DC109306 (05/1/107-NR)

Batch code: 11565
Exp. Date: 12/2017
Manufactured: 09/2015
Tolerance: +/- 0.02 pH

Nominal pH-value: 9.18 pH (at 25°C)
Actual pH-value: 9.18 pH (at 25°C)
Secondary Ref. Material: Sodiumtetraborate

Ref. Material LOT-No.: A0264508321

Standardization method:

Actual pH-value as stated above was obtained using a combination glass electrode according to DIN 19268.

Calibration is traceable to national and international standards, PBT and NIST validated by key comparisons.

WATER-1.D®
PRODUCTION BIMBH

20 Delinforstreese 20
D-76344 Egyandsin-Leoph.(6=xxxy)

3 +49 (0) 721 - 78 20 29 - 0

48 (3) 721 - 78 20 29 - 11

WATER-1.D®

Date: Sep. 21st, 2015

Eggenstein, Germany

em And

Head of Lab