

CERTIFICATE OF ANALYSIS

Buffer Solution

1.68 pH

Item code:	DC101680 (05/1/107-NR)
Batch code:	11638
Exp. Date:	11/2017
Manufactured:	11/2015
Tolerance:	+/- 0.015 pH
Nominal pH-value:	1.68 pH (at 25°C)
Actual pH-value:	1.68 pH (at 25°C)
Secondary Ref. Material:	Potassiumtetraoxalate
Ref. Material LOT-No.:	M1680S01/14

Standardization method:

Actual pH-value as stated above was obtained using a combination glass electrode according to DIN 19268.

Calibration is traceable to national and international standards, PBT and NIST validated by key comparisons.