

CERTIFICATE OF ANALYSIS

Buffer Solution

10.01 pH

Item code:	DC110078 (05/1/107-NR)
Batch code:	12048
Exp. Date:	08/2020
Manufactured:	05/2018
Tolerance:	+/- 0.03 pH
Nominal pH-value:	10.01 pH (at 25°C)
Actual pH-value:	10.01 pH (at 25°C)
Secondary Ref. Material:	Sodiumtetraborate
Ref. Material LOT-No.:	A0264508321

Standardization method:

Actual pH-value as stated above was obtained using a combination glass electrode according to DIN 19268.

Calibration is traceable to national and international standards, PBT and NIST validated by key comparisons.

WATER-I.D.[®]
POOL-I.D. PRIMELAB EASYDIP
Daimlerstrasse 20
D-76344 Eggenstein-Leoph.(Germany)
+49 (0) 721 - 78 20 29 - 0
+49 (0) 721 - 78 20 29 - 11
www.water-id.com • info@water-id.com

Date: MAY 6th, 2018

Eggenstein, Germany



Head of Lab