

## CERTIFICATE OF ANALYSIS

# Buffer Solution

# 10.01 pH

Item code:	DC110078 (05/1/107-NR)
Batch code:	11501
Exp. Date:	08/2017
Manufactured:	08/2015
Tolerance:	+/- 0.03 pH
Nominal pH-value:	10.01 pH (at 25°C)
Actual pH-value:	10.01 pH (at 25°C)
Secondary Ref. Material:	Sodiumtetraborate
Ref. Material LOT-No.:	A0264508321

### Standardization method:

Actual pH-value as stated above was obtained using a combination glass electrode according to DIN 19268.

Calibration is traceable to national and international standards, PBT and NIST validated by key comparisons.